PLAINTIFF'S EXHIBIT

UC-2734

Asbestos - Institute

Analyzed
~~Calculated~~ 8-21-80
by Ruth Close (DRC)

HHF-113

21 fibers/100 fields
.21 fibers/field

$$\frac{.21 \times 855}{6.24 \times 30} = 1.0 \text{ fiber} > 5\mu m/cm^3$$

HHF-114

11 fibers/100 fields
.11 fibers/field

$$\frac{.11 \times 855}{6.24 \times 30} = .5 \text{ fiber} > 5\mu m/cm^3$$

HHF-115

14 fibers/100 fields
.14 fiber/field

$$\frac{.14 \times 855}{6.24 \times 30} = .6 \text{ fibers} > 5\mu m/cm^3$$

Union Carbide trial exhibit UC-2734: Handwritten Fiber-Count Calculations for Asbestos Institute Samples HHF-113... Published by AsbestosClaims.Law
ACL Doc 17645 · Page 1 of 2 **asbestosclaims.law/library/17645/1**

HHF-116

2 fibers/100 fields

.02 fibers/field

$\frac{.02 \times 855}{6.24 \times 34} =$ < .1 fiber > 5 μm/cm³

HHF-116

108 fibers/12 fields

200 fibers/20 fields

10 fibers/field

$\frac{10 \times 855}{6.24 \times 30} =$ 45.7 fibers > 5 μm/cm³

UCC 024651